



NOTES

1. RESISTORS R11 AND R12 ARE A.C. VOLTAGE CALIBRATION RESISTORS WHOSE VALUES ARE DETERMINED BY THE RECTIFIER CHARACTERISTICS.
2. RESISTORS R29, R21 AND R22 ARE THE HIGH END, LOW END AND LINE VOLTAGE CALIBRATORS RESPECTIVELY WHOSE VALUES ARE DETERMINED DURING THE CALIBRATION OF THE INSTRUMENT.

ANALYZER SECTION

TUBE TESTER SECTION

STARK MODEL 9-54T SERVICE-MASTER

STARK ELECTRONIC INSTRUMENTS LTD  
 TORONTO CANADA

9-54T SCHEMATIC

ISSUE 1 OCT 29 1946	PART NUMBER
D'W N CH'Y'D APPROVED	
<i>[Signature]</i>	